

gloveon

Innova Polymer Coated

Latex Exam Gloves Powder Free, Standard Cuff



The **Innova Polymer Coated** powder free glove dons like a powdered glove but without the concern related to the potential adverse health effects from using one. This glove offers softness and flexibility, and due to the excellent barrier properties, it provides high tensile strength and dexterity for the user. This glove is widely used in laboratory environments and critical healthcare environments where prevention of contamination is of utmost importance.



GloveOn Innova Polymer Coated

GloveOn Innova Polymer Coated		
Length (mm)	≥ 230	
Thickness Measurements (mm)		
Palm (centre of Palm)	0.14 ± 0.03	
Finger (13mm \pm 3mm from tip)	0.15 ± 0.03	
Physical Properties	Before Ageing	After Ageing
Tensile Strength (MPa)	≥ 18	≥ 16
Elongation (%)	≥ 650	≥ 500
Inspection Levels & AQL	Inspection Level	AQL
Watertightness	G1	1.50
Physical Dimensions	S2	4.00
Tensile Strength	S2	4.00
Visual Inspection (Major)	S4	2.50
Visual Inspection (Minor)	S4	4.00
Particulate Residue	N = 5	$\leq 2\text{mg/glove}$
Protein Content	N = 3	$\leq 50\mu\text{g/g}$

REORDER CODE

PLM22XS	X-SMALL
PLM22SS	SMALL
PLM22MM	MEDIUM
PLM22LL	LARGE
PLM22XL	X-LARGE

FEATURES

- Fully textured
- Polymer Coated Powder free
- Made with natural rubber latex
- Protein level $\leq 50\mu\text{g/g}$
- Ambidextrous
- Standard cuff
- Natural colour

PACKAGING

100 gloves per box for XS to XL
10 boxes per carton

REGULATORY COMPLIANCE

TGA - ARTG 127794, FDA 510(k),
MDD 93/42/EEC, REACH,
CMDR SOR-98-282

STANDARDS

ASTM D3578, EN 455 part 1, 2, 3 & 4,
ISO 11193 - 1, AS/NZS 4011:1997

MANUFACTURING ACCREDITATIONS

ISO 9001:2008
ISO 13485:2003
EN ISO 13485:2003

*Full test reports on Sensitisation Testing, Primary Skin Irritation and Residual Proteins Testing are available upon request.

A brand by



Mun Australia
Toll free: 1800 456 837 (GLOVES)
info.au@munglobal.com



Protection Always On