

Clean-up of Sequencing Reactions

User Manual

NucleoSEQ®

December 2010/Rev.03

Table of contents

1	Components	4
1.1	Kit contents	4
1.2	Material to be supplied by user	4
2	Product description	5
2.1	The basic principle	5
2.2	Kit specifications	5
3	Storage conditions	5
4	Safety instructions – risk and safety phrases	5
5	Protocol for sequencing reaction clean-up	6
6	Appendix	8
6.1	Troubleshooting	8
6.2	Ordering information	8
6.3	Product use restriction/warranty	9

1 Components

1.1 Kit contents

NucleoSEQ®			
	10 preps	50 preps	250 preps
REF	740523.10	740523.50	740523.250
NucleoSEQ® Columns	10	50	250
Collection Tubes	10	50	250
User Manual	1	1	1

1.2 Material to be supplied by user

- Collection tubes (e.g., 1.5 mL microcentrifuge tubes)

2 Product description

2.1 The basic principle

NucleoSEQ® Columns are designed for fast and effective clean-up of nucleic acids. Using gel exclusion in a convenient spin column format allows reliable removal of smaller molecules from nucleic acids. Impurities, for example, salts, excess of labels, nucleotides, traces of organic solvents, primers are retained by the column while nucleic acids of interest are recovered with high yield. The columns are pre-filled with size exclusion matrix.

2.2 Kit specifications

- Maximum sample volume to be loaded onto the column: 20 µL
- Removal of sequencing dye terminators including BigDye®

3 Storage conditions

- NucleoSEQ® Columns with dry gel matrix can be stored at room temperature for 12 months.
- NucleoSEQ® Columns with hydrated gel matrix should be stored at 4°C. Columns can be stored up to 14 days at 4°C.

4 Safety instructions – risk and safety phrases

The components of the NucleoSEQ® kits do not contain hazardous contents.

5 Protocol for sequencing reaction clean-up

General procedure

(For details on each step see page 7.)

- 1 Spin down dried gel resin



750 x g
30 s

- 2 Hydrate gel resin with **600 µL water**, vortex, and incubate at least **30 min** for complete hydration



600 µL
water


RT
> 30 min

- 3 Remove bottom plug and spin down hydrated gel resin



750 x g
2 min

- 4 Load sample to the center of the column



Load
column

- 5 Spin for **4–6 min** at **750 x g** to recover purified sample



750 x g
4 – 6 min

Detailed procedure

Perform sequencing reaction according to standard protocols.

We recommend to use not more than 1–2 µL of Big Dye® Ready reaction mix in a 20 µL sequencing reaction in order to avoid overloading the column.

- 1** Centrifuge the NucleoSEQ® Columns for **30 s** at **750 x g** to collect the dry gel matrix on the bottom of the cartridge.

- 2** Add **600 µL dist. water** and vortex to hydrate the gel matrix. Remove air bubbles by vortexing or tapping the column. Incubate at least **30 min** or overnight to hydrate the gel matrix. Incubation can be performed at room temperature or 4°C. Hydrated columns can be stored at 4°C for a maximum of 14 days. Resuspend the settled gel matrix by inverting or vortexing the spin column several times. Air bubbles should not be visible now. Remove the bottom plug and place the spin column into a Collection Tube (supplied with the kit).

- 3** Place the column into an appropriate centrifuge (the hinge of the spin column's cover lid should be orientated to the outside of the rotor). Centrifuge **2 min** at **750 x g** to remove the remaining storage buffer. Discard the collection tube with storage buffer. Place the spin column in an appropriate collection tube (e.g., microcentrifuge tube, not supplied with the kit).

- 4** Open the lid of the column. Carefully load the sample drop by drop onto the center of the gel resin. Pipetting the sample at the sides of the spin column tube may reduce purification efficiency of the column. Moreover, do not disturb the gel surface. Sample volume should not exceed 20 µL.

- 5** Place the column in the same orientation as in step 3 into the centrifuge. Elute the sample by centrifuging the column for **4–6 min** at **750 x g**. Discard the spin column.

Dry the sample or use the sample directly.

6 Appendix

6.1 Troubleshooting

Problem	Possible cause and suggestions
Dye blobs	<i>Sample added to column improperly</i> <ul style="list-style-type: none"> Add sample directly to the center of the settled gel matrix. Dispense sample dropwise. Avoid adding sample at the sides of the gel matrix
	<i>Sample volume to high</i> <ul style="list-style-type: none"> Add sample in a volume of 20 µL. Higher sample volumes can cause incomplete removal of dye terminators.
Poor signal intensity	<i>Sample volume to small</i> <ul style="list-style-type: none"> Add sample in a volume of 20 µL. If necessary adjust sample volume to 20 µL using distilled water

Conversion of RCF from different centrifuges:

$$\text{rpm} = 100 \times \sqrt{\text{RCF} / 1.12r}$$

$$\text{RCF} = (\text{rpm} / 1000)^2 \times 1.12r$$

rpm = revolutions per minute

RCF = relative centrifugal force (g force = RCF x g)

r = radius in mm

6.2 Ordering information

Product	REF	Pack of
NucleoSEQ®	740523.10	10 preps
	740523.50	50 preps
	740523.250	250 preps

Visit www.mn-net.com for more detailed product information.

6.3 Product use restriction/warranty

NucleoSEQ® kit components are intended, developed, designed, and sold FOR RESEARCH PURPOSES ONLY, except, however, any other function of the product being expressly described in original MACHEREY-NAGEL product leaflets.

MACHEREY-NAGEL products are intended for GENERAL LABORATORY USE ONLY! MACHEREY-NAGEL products are suited for QUALIFIED PERSONNEL ONLY! MACHEREY-NAGEL products shall in any event only be used wearing adequate PROTECTIVE CLOTHING. For detailed information please refer to the respective Material Safety Data Sheet of the product! MACHEREY-NAGEL products shall exclusively be used in an ADEQUATE TEST ENVIRONMENT. MACHEREY-NAGEL does not assume any responsibility for damages due to improper application of our products in other fields of application. Application on the human body is STRICTLY FORBIDDEN. The respective user is liable for any and all damages resulting from such application.

DNA/RNA/PROTEIN purification products of MACHEREY-NAGEL are suitable for *IN-VITRO*-USES ONLY!

ONLY MACHEREY-NAGEL products specially labeled as IVD are also suitable for *IN-VITRO*-diagnostic use. Please pay attention to the package of the product. *IN-VITRO*-diagnostic products are expressly marked as IVD on the packaging.

IF THERE IS NO IVD SIGN, THE PRODUCT SHALL NOT BE SUITABLE FOR *IN-VITRO*-DIAGNOSTIC USE!

ALL OTHER PRODUCTS NOT LABELED AS IVD ARE NOT SUITED FOR ANY CLINICAL USE (INCLUDING, BUT NOT LIMITED TO DIAGNOSTIC, THERAPEUTIC AND/OR PROGNOSTIC USE).

No claim or representations is intended for its use to identify any specific organism or for clinical use (included, but not limited to diagnostic, prognostic, therapeutic, or blood banking). It is rather in the responsibility of the user or - in any case of resale of the products - in the responsibility of the reseller to inspect and assure the use of the DNA/RNA/protein purification products of MACHEREY-NAGEL for a well-defined and specific application.

MACHEREY-NAGEL shall only be responsible for the product specifications and the performance range of MN products according to the specifications of in-house quality control, product documentation and marketing material.

This MACHEREY-NAGEL product is shipped with documentation stating specifications and other technical information. MACHEREY-NAGEL warrants to meet the stated specifications. MACHEREY-NAGEL's sole obligation and the customer's sole remedy is limited to replacement of products free of charge in the event products fail to perform as warranted. Supplementary reference is made to the general business terms and conditions of MACHEREY-NAGEL, which are printed on the price list. Please contact us if you wish to get an extra copy.

There is no warranty for and MACHEREY-NAGEL is not liable for damages or defects arising in shipping and handling (transport insurance for customers excluded), or out of accident or improper or abnormal use of this product; defects in products or com-

ponents not manufactured by MACHEREY-NAGEL, or damages resulting from such non-MACHEREY-NAGEL components or products.

MACHEREY-NAGEL makes no other warranty of any kind whatsoever, and SPECIFICALLY DISCLAIMS AND EXCLUDES ALL OTHER WARRANTIES OF ANY KIND OR NATURE WHATSOEVER, DIRECTLY OR INDIRECTLY, EXPRESS OR IMPLIED, INCLUDING, WITHOUT LIMITATION, AS TO THE SUITABILITY, REPRODUCTIVITY, DURABILITY, FITNESS FOR A PARTICULAR PURPOSE OR USE, MERCHANTABILITY, CONDITION, OR ANY OTHER MATTER WITH RESPECT TO MACHEREY-NAGEL PRODUCTS.

In no event shall MACHEREY-NAGEL be liable for claims for any other damages, whether direct, indirect, incidental, compensatory, foreseeable, consequential, or special (including but not limited to loss of use, revenue or profit), whether based upon warranty, contract, tort (including negligence) or strict liability arising in connection with the sale or the failure of MACHEREY-NAGEL products to perform in accordance with the stated specifications. This warranty is exclusive and MACHEREY-NAGEL makes no other warranty expressed or implied.

The warranty provided herein and the data, specifications and descriptions of this MACHEREY-NAGEL product appearing in MACHEREY-NAGEL published catalogues and product literature are MACHEREY-NAGEL's sole representations concerning the product and warranty. No other statements or representations, written or oral, by MACHEREY-NAGEL's employees, agent or representatives, except written statements signed by a duly authorized officer of MACHEREY-NAGEL are authorized; they should not be relied upon by the customer and are not a part of the contract of sale or of this warranty.

Product claims are subject to change. Therefore please contact our Technical Service Team for the most up-to-date information on MACHEREY-NAGEL products. You may also contact your local distributor for general scientific information. Applications mentioned in MACHEREY-NAGEL literature are provided for informational purposes only. MACHEREY-NAGEL does not warrant that all applications have been tested in MACHEREY-NAGEL laboratories using MACHEREY-NAGEL products. MACHEREY-NAGEL does not warrant the correctness of any of those applications.

Last updated: 07/2010, Rev. 03

Please contact:
MACHEREY-NAGEL GmbH & Co. KG
Tel.: +49 (0) 24 21 969 270
e-mail: tech-bio@mn-net.com

Trademarks/Disclaimer:

BigDye® is a trademark of a member of Applied Biosystems
NucleoSEQ® is a registered trademark of MACHEREY-NAGEL GmbH & Co. KG

All used names and denotations can be brands, trademarks, or registered labels of their respective owner – also if they are not special denotation. To mention products and brands is only a kind of information (i.e., it does not offend against trademarks and brands and can not be seen as a kind of recommendation or assessment). Regarding these products or services we can not grant any guarantees regarding selection, efficiency, or operation.